

AMDM 2015 Focus Meeting



Update on Japanese *In Vitro* Diagnostics Regulations

Motoki MIKAMI

*Reviewer, Office of In Vitro Diagnostics
Pharmaceuticals and Medical Devices Agency
Tokyo, Japan*



Agenda

- I. Overview of Japanese IVD Regulation
- II. Revised Points of
Pharmaceuticals and Medical Devices Act
- III. Current Topics

Agenda

- I. Overview of Japanese IVD Regulation
- II. Revised Points of
Pharmaceutical and Medical Device Act
- III. Current Topics

Role of MHLW and PMDA

- **Ministry of Health, Labor and Welfare (MHLW)**

Planning basic policy, enforcement of administrative measures based on the law

- Ex. ● Final judgment on approval of pharmaceuticals and medical devices
- Issue emergency safety information and direct product withdrawal from market

- **Pharmaceuticals and Medical Devices Agency (PMDA)**

Review, examination and data analysis

- Ex. ● Scientific review of pharmaceuticals and medical for marketing authorization
- Consultation on the development of pharmaceuticals and medical devices
 - GMP/GLP/GCP inspections
 - Collection and analysis of adverse Event Reports



IVD reagents and IVD equipments (1)

- Pharmaceuticals and Medical Devices Act (PMD Act)
: Article 2 Definitions

Medical Devices

- devices to be used for the **diagnosis**, treatment or prevention of humans or animals, to effect on the body structure or function of humans or animals (Clause 4)

Drugs

- Substances (other than quasi-drugs) which are intended for use in the **diagnosis**, treatment or prevention of diseases in humans or animals, and which are not equipment, dental materials, medical supplies, sanitary materials or programs (Clause 1 ii)

In vitro diagnostic reagents (drugs)


- **drugs** which are intended for use in the diagnosis of diseases, and which are not used directly on humans or animals (Clause 14)

IVD reagents and IVD equipments (2)

➤ Under the Pharmaceuticals and Medical Devices Act (PMD Act)

* In vitro diagnostics **reagents(drugs)** are categorized as **drugs**.

* In vitro diagnostics **equipments** are categorized as **medical devices**.



➤ Although IVD reagents are categorized as drugs, IVD reagents are controlled under the similar rules as the medical devices.

- Risk based classification
- QMS requirements
- Essential principle
 - Generic name etc.

Classification of IVD reagents

Risk	Class	Risk Level and Examples
Low Risk	Class I	<p>Its diagnostic information risk is relatively low when used in the diagnosis of disease and which its accuracy in terms of the information which they provide is believed to have a relatively weak impact on life support in comparison with Class III Products. These products are recognized by ministerial notification that they have standards and/or standardized measuring method and that they can be calibrated relatively easily upon manufacturing/quality management. OTC products are not included.</p> <p>Example : GOT, GPT, Glucose, LDH, Estradiol</p>
	Class II	<p>Its diagnostic information risk is relatively low when used in the diagnosis of disease and which its accuracy in terms of the information which they provide is believed to have a relatively weak impact on life support in comparison with Class III Products. OTC products are included.</p> <p>Example : Blood Cell Morphology, Autoimmunity</p>
Other	Class III	<p>Its diagnostic information risk is relatively high when used in the diagnosis of disease and which its accuracy in terms of the information which they provide is believed to have a relatively strong impact on life support.</p> <p>Example : HIV, HCV, Tumor markers, Microbiology</p>

Type of IVD reagents Regulation

Regulation	Category
Self Certification	Class I products with self-certification standard
Certification by Registered Certification Body	Class II products with certification standard
Approval by MHLW (PMDA's review)	Class III products, Class I / II products not conforming standard, Novel products

Number of Approvals and Consultations

Number of approvals and review times for IVD


Fiscal Year	FY2010	FY2011	FY2012	FY2013	FY2014
Number of Approvals	191	173	147	166	109
Total Review Times [months]	8.2	7.4	6.0	5.4	5.3
Regulatory Review Times [months]	5.8	4.1	3.4	2.7	2.6

Number of consultations conducted for IVD

Fiscal Year	FY2010	FY2011	FY2012	FY2013	FY2014
Number of Consultations	7	5	8	7	25

*Source: PMDA annual report FY2014

10



Collaboration plan to accelerate review of IVD

(issued on 31 March 2014)

This plan from FY 2014 to FY 2018 aims to shorten and standardize the review periods of IVD approval through the collaboration of Japanese regulatory authorities and industries by;

1. Taking measures to improve the quality of review process

Training for applicants, Efficient consultation service, Standardization of review process

2. Setting standard review periods

80th percentile value out of application cohorts of IVD reagents should satisfy each targeted overall review period from application to approval. (**13 months** for IVD with consultation with external experts and **7 months** for normal IVD)

3. Increase of reviewers

4. Managing progress of this plan

Regulatory authorities and industries will check its progress regularly and take necessary action(s) to ensure proper implementation.

Agenda

- I. Overview of Japanese IVD Regulation
- II. Revised Points of
Pharmaceutical and Medical Device Act**
- III. Current Topics

Background on Revision of Pharmaceutical Affairs Law

- High demands to market innovative drugs and medical devices as well as advance R&D of cellular and tissue therapeutic products in a timely manner
- Needs for stronger safety measures
- The Japan Revitalization Strategy

Brief Overview of Revision of PAL

- Points of this revision are to;
 1. Strengthen safety measures regarding drugs and medical devices
 2. Revise medical device regulations based on its characteristics
 3. Introduce cellular and tissue therapeutic product regulations based on its characteristics
- The law has been renamed as “Act on Securing Quality, Efficacy and Safety of Pharmaceuticals, Medical Devices, Regenerative and Cellular Therapy Products, Gene Therapy Products, and Cosmetics” = “**PMD Act**”.
- The chapter for medical devices/IVDs is prepared separately.

→ PMD Act has been put into effect on 25 November 2014


Summary of Main Revisions Related to IVD Reagents

1. Marketing Authorization Holder (MAH) of IVD reagents is re-defined.
(It **was not distinguished from MAH of drug** under PAL.)
2. **Manufacturer** of IVD reagents should be **registered**, not licensed or accredited.
3. QMS inspection is performed **per product family**, not on individual product.
Product family of IVD reagents are categorized in the following;
A)Radioactive IVD reagents
B)Other IVD reagents than those in A)
4. Pre-marketing **application form** and its **attachments** are revised.


Scope of Facilities to be Registered as a Manufacturer

Manufacturing Process	Radioactive IVD reagents	Non-radioactive IVD reagents
Design	X	X
Filling process of reagents involved in a reaction system in final product	X	X
After filling process of reagents before storage of final product	X	NA
Domestic storage site of final product	X	X

Example of Registration as a Manufacturer



QMS Inspection is applied to MAH, not each Manufacturer



19

Agenda

- I. Overview of Japanese IVD Regulation
- II. Revised Points of
Pharmaceuticals and Medical Devices Act
- III. Current Topics**

Definition of CoDx in Japan(1)

A companion diagnostics(CoDx) is essential for using the pertinent therapeutic product, and corresponds to either of the following (*except *in vitro* diagnostic agents or medical devices intended simply for disease diagnosis, etc.*) :

- that is used to identify patients who are expected to *respond better* to a specific therapeutic product.
- that is used to identify patients who are likely to be at *high risk of developing adverse events* associated with a particular therapeutic product
- that is necessary for *optimizing the treatment* including dose, schedule, and discontinuation of a particular therapeutic product

[July 2013; Notification on Approval Application for *In Vitro* Companion Diagnostics and Corresponding Therapeutic Products](#)

21

Definition of CoDx in Japan(2)

Q: What examples are there of “*in vitro* diagnostic agents or medical devices intended simply for disease diagnosis” described in Section 1 of this Notification?

A: Examples may include *in vitro* diagnostics that are used for
biochemical assays related to organ functions such as serum creatinine, transaminases, and blood glucose level,
hematological assays such as prothrombin time kit,
bacterial or viral identification and susceptibility tests for infections,
as well as **tests used to identify the disease**, check the treatment effect, assist in follow-up observation, or evaluate the severity **in routine clinical practice**.
However, diagnostics of these types may also be judged as CoDx depending on the clinical necessity, etc.

22

Discussion Points of NGS Regulation

- Is each component of NGS **MD** or **IVD reagent** in Japan?
- How to evaluate **analytical performance** of NGS
- How to **define validated mutations** as indication of corresponding drugs
- How to deal with **update of clinical database** after marketing authorization of NGS (and corresponding drug)

Companion Diagnostics WG

Companion diagnostics (CoDx) working group in PMDA

One of the projects across multi-offices in PMDA

Founded in April, 2012

HP: <http://www.pmda.go.jp/rs-std-jp/standards-development/cross-sectional-project/0013.html>

Common Mistakes Resulting in the Delay of IVD Approval(1)

- Stability Test for Shelf Life

Requirements are specific in Japan;

- Method : **QC release test** and the criteria described in application form
- Lots : **3 lots**
- Repeats : **2** or more
- Time points : **3 times or more**; **Beginning, Interim, Claimed shelf life or later** in real time
- Storage condition : **Equal** or **above the upper limit** of the claimed temperature
(This shall not apply in case of refrigerated.)
- * Shipping stability, In use stability; self-assurance

- Clinical Performance Study for Novel Product

Sometimes submitted protocol and clinical data are insufficient in followings;

- Generalisability
- Positioning of the product
- Difference from previous product
- Medical environment difference from foreign countries etc.

Common Mistakes Resulting in the Delay of IVD Approval(2)

- Clinical Performance Study for Companion Diagnostics (CoDx)

1) **CoDx-negative cases** (i.e. patients not enrolled in clinical trial) **also** should be evaluated in clinical performance study.


- Confirming the test was conducted appropriately
- Validating true negative (not false negative) etc.

2) In some cases, CoDx testing results of **exploratory clinical trials** are required.
(e.g. if clinical cut off is defined in the **exploratory clinical trial**)

Take home message

To further shorten the review period, we need ••

- **Applicant's cooperation**
 - appropriate evaluation of the IVDs
 - suitably-preparation of application documents
- **Oversea manufacturer's cooperation**
 - giving accurate information to Japanese distributors
 - quick response to inquiry from Japan
- **Joining the meeting with PMDA (as necessary)**



27

Thank you very much
for your attention !



<http://www.pmda.go.jp/>

28