




The FDA PMA Process for *In Vitro* Diagnostic Devices

FDA/Center for Devices and Radiologic
Health/Office of *In Vitro* Diagnostics

April 21th 2010



When is a PMA Necessary?

Presented by
Kate Simon, PhD


FDA Regulated Uses of IVDs

- **Diagnosis** – Diagnose disease, identify pathogens, confirm, or rule out infection in symptomatic patients
- **Screening** - Intended use population includes individuals **without** signs or symptoms of disease, infection
- **Epidemiology/Surveillance** - To detect and monitor incidence or prevalence of infection for targeting and evaluating health programs
- **Monitoring, prognosis, prediction**

**I have a device with one of these intended uses
what kind of FDA submission should I prepare ?**

How are IVD Devices Classified?

Class I – most 510(k)
exempt



Low likelihood
of harm

Class II - 510(k)

High or unknown
likelihood of
harm,

or how to prevent
harm is unknown

Class III - PMA

- **Regulatory path** determined using a risk-based approach
- **Classification** (I, II, or III) depends on risk



Risk is Dependent Upon Intended Use

- Risk (and subsequently classification and submission type) is inherently tied to **Intended Use** of a device.



Risk is Dependent Upon Intended Use

- Level of FDA review and type of studies requested generally depend on the Intended Use claims; not always on type of technology or assay
- Prostate-specific antigen (PSA) testing with an indication for
 - cancer screening & diagnosis (PMA)
 - prognosis & monitoring (510(k))

Risk is Dependent Upon Intended Use

- A CFTR genotyping assay with the indication
 - ✓ For aid in diagnosis →**510(k)**
 - ✓ For fetal screening →**PMA**
- One multiplex instrument system with 2 devices
 - ✓ Device detecting BCR-ABL for CML diagnosis →**PMA**
 - ✓ Device detecting BCR-ABL for monitoring →**510(k)**



For Established IVD Devices

- Search our Classification Database to determine device class and required submission type:

<http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpcd/classification.cfm>



For Novel IVD Devices

- Can the device be placed under existing regulations?
- If not, then the classification and submission type must be determined

When is a Device Class III?

- Class III devices are those:
 - a) that cannot be classified as class II because insufficient information exists to determine that special controls would provide reasonable assurance of its safety and effectiveness;
 - b) that cannot be classified as class I because "insufficient information exists to determine that the application of general controls [is] sufficient to provide reasonable assurance of safety and effectiveness of the device";

AND...

When Class III ? cont...

- c) and that "(I) is purported or represented to be for a use in supporting or sustaining human life or for a use which is of substantial importance in preventing impairment of human health,
- d) or (II) presents a potential unreasonable risk of illness or injury." Section 513(a)(1)(C) (21 U.S.C. 360c(a)(1)(C)).



Before FDA Modernization Act

- 513 (f)(1) of F, D, & C Act automatically classifies devices that were not in commercial distribution prior to May 28, 1976 into Class III, requiring a pre-market approval (PMA)



FDA Modernization Act of 1997

- Provides a new mechanism for classifying new devices for which there is no predicate device
- Allows an automatic class III designation to be evaluated and overturned
- We call this mechanism the De Novo Process

FDA Modernization Act of 1997 (FDAMA) - New Section 513(f)(2) of the F, D, & C Act. Amended November 21, 1997



Downclassification of Class III Devices

- Class III devices can be downclassified to Class II when sufficient information becomes available to establish **special controls** that reasonably assure safety and effectiveness

Downclassification of Existing Class III Devices

- Downclassification of an existing Class III device - citizen's petition
- **Recent example:** Hepatitis A infection diagnostic devices. **Reassessment of level of risk**
- Hepatitis B and C infection diagnostic devices remain as Class III



Other Regulatory Tools

- **513g** – Official request for classification of a currently unclassified device
- **Pre-IDE submission** – Informal interactive process allowing early assessment of device class, and least burdensome regulatory route to approved product



Comparison of the PMA and 510(k) Processes

Presented by
Zivana Tezak, PhD



Outline

- Terminology
- Elements - PMA or 510(k)
 - Intended use
 - PMA specific sections
 - Analytical performance
 - Clinical performance
 - Labeling

Terminology

Class	Pre-market Submission	Success Metric	Action
III	PMA	Safety and Effectiveness	Approval
II	510(k)	Substantial Equivalence	Clearance
I	None (if exempt)		
II (De Novo)	510(k)	Safety and Effectiveness	Clearance

Class III Devices

- Regulation governing premarket approval
 - in Title 21 CFR Part 814
- Act Section 515 (d)(6):
 - PMA supplements required for changes affecting safety and effectiveness
 - For manufacturing changes - a 30-day notice or 135-day PMA supplement
- Timeline - FDA has 180 days to review the PMA and make a determination

Major Elements of an IVD Submission

- Intended use/indications for use
- Device description, internal / external controls
- Pre-analytical (e.g. sample prep) and analytical performance
- Clinical performance
- Instrument and software, if applicable
 - If multiple platforms, assay performance on each
- Labeling (package insert) - "truth in labeling"

Intended Use

What assay measures, how to use results

*Intended
Population*

Analyte

*Indication
For Use*

Example:

*MammaPrint® is a qualitative in vitro diagnostic test service performed in a single laboratory, using the **gene expression profile** of fresh frozen breast cancer tissue samples to assess a patients' risk for distant metastasis.*

*The test is performed for **breast cancer patients** who are less than 61 years old, with Stage I or Stage II disease, with tumor size ≤ 5.0 cm and who are lymph node negative. The MammaPrint® result is indicated for use **by physicians as a prognostic marker only, along with other clinicopathological factors.***

Types of **studies** depend on IU claims;
may also depend on the technology or assay format



PMA Specific Elements

- Manufacturing section
- Pre-approval inspection (GMP compliance)
- BIMO (bioresearch monitoring visit to clinical and/or sponsor sites)
- Possible Panel-Track (novel IU)
- Post-approval – annual reports, PMA supplements for well defined modifications
<http://www.fda.gov/cdrh/ode/guidance/1584.pdf>

Analytical Validation – Quality of Measurement

- **Analytical performance measures**
 - Precision (repeatability, reproducibility)
 - Accuracy
 - Sensitivity, Limit of Detection
 - Specificity (interference, cross-reactivity)
 - Sample type / matrix
 - Sample preparation / conditions
 - Performance around the cut-off
 - Potential for carryover, cross-hybridization
 - Stability (for PMA)

- **Studies may vary depending on:**
 - Technology, end user
 - Quantitative or qualitative assay
 - What is reported (individual analytes vs. composite score)



Performance

- **Analytical performance**—does my test measure the analyte I think it does? Correctly? How reliably?
- **Clinical performance**—does my test result correlate with the expected clinical presentation? How reliably?

Accuracy / Clinical Performance

- Real clinical samples where feasible
- Prospective or retrospective evaluation
- Comparison to a **reference method**
e.g., bi-directional DNA sequencing for genotyping; viral culture; composite methods
- Comparison to a **predicate device**
 - 510(k)
- Comparison to a **clinical outcome**
 - PMA, but also
 - Some 510(k)s & de novo 510(k)s



Clinical Validation – Significance of Result

- Study plan for an *in vitro* diagnostic product depends on the intended use / indications for use/end user
- Diagnosis, residual disease, etc. (current state)
- Monitoring, recurrence (change in state)
- Risk of disease, prognosis, prediction (future state)



Clinical Section of a PMA Submission

- Study protocols including IRB approval letters/informed consent
- Safety and effectiveness data
- Adverse reactions and complications
- Device failures and replacements
- Case report forms, patient information, patient complaints, any studies done under IDE
- Tabulations of data from all individual subjects
- Data analysis, results of statistical analyses
- Any other information from the clinical investigations
- Literature




Labeling of *In Vitro* Devices

- 21 CFR 809.10
- Clear instructions for use
- Need to capture expected analytical and clinical performance of device
- Prospective performance in intended use population

Approval Documents

- PMA approval - summary of the safety and effectiveness data upon which the approval is based, labeling available
(<http://www.fda.gov/cdrh/pmapage.html#monthly>)



The screenshot displays the FDA's Center for Devices and Radiological Health (CDRH) website. At the top, the FDA logo and "U.S. Food and Drug Administration" are visible, along with the Department of Health and Human Services logo. Below this, the CDRH name and a navigation bar with links like "FDA Home Page", "CDRH Home Page", "Search", and "A-Z Index" are present. A "Questions?" button is also located in the top right. The main content area shows the breadcrumb trail: "FDA > CDRH > Information on Premarket Approval Applications > COBAS® TaqMan® HBV Test For Use With the High Pure System - P050028". The title "COBAS® TaqMan® HBV Test For Use With the High Pure System - P050028" is prominently displayed. Below the title, it states "Issued September 4, 2008". A list of links is provided: "Approval Order", "Summary", "Labeling", and "Other Consumer Information". In the bottom right corner, it says "Updated October 2, 2008". The footer contains a detailed navigation bar with links such as "CDRH Home Page", "CDRH A-Z Index", "Contact CDRH", "Accessibility", "Disclaimer", "FDA Home Page", "Search FDA Site", "FDA A-Z Index", "Contact FDA", and "HHS Home Page". At the very bottom, it reads "Center for Devices and Radiological Health / CDRH".

FDA U.S. Food and Drug Administration Department of Health and Human Services
CENTER FOR DEVICES AND RADIOLOGICAL HEALTH
[FDA Home Page](#) | [CDRH Home Page](#) | [Search](#) | [A-Z Index](#) [Questions?](#)

[FDA](#) > [CDRH](#) > [Information on Premarket Approval Applications](#) > COBAS® TaqMan® HBV Test For Use With the High Pure System - P050028

COBAS® TaqMan® HBV Test For Use With the High Pure System - P050028

Issued September 4, 2008

- [Approval Order](#)
- [Summary](#)
- [Labeling](#)
- [Other Consumer Information](#)

Updated October 2, 2008

[CDRH Home Page](#) | [CDRH A-Z Index](#) | [Contact CDRH](#) | [Accessibility](#) | [Disclaimer](#)
[FDA Home Page](#) | [Search FDA Site](#) | [FDA A-Z Index](#) | [Contact FDA](#) | [HHS Home Page](#)

Center for Devices and Radiological Health / CDRH



Some Common Questions

- Are clinical studies for a PMA always more extensive than for a 510(k)?

Not Always

- When to register and list?

<http://www.fda.gov/cdrh/registration/whento.html>

- What is available to streamline the process?



How to Avoid Potential Pitfalls in the PMA Process

Presented by
Sally Hojvat, PhD



Outline

- Reasons why the PMA submission review/approval process may take longer than you expected
- How to improve your PMA submission
- Ways to streamline the PMA approval process

FDA PMA Review Oversight

- This is how a PMA arrives to our Office !

PMA Team Formed

- Lead reviewer
- Statistician
- Compliance
- Epidemiologist
- Internal/external experts in field
- Instrument/software expert etc.



Review/Approval Takes Longer than Expected. Why? (1)

Global Issues with Submissions

○ Disorganized

Table of contents missing, pages not numbered

- Check tables/figs./text for clarity, consistency and accuracy
- “Put together in a hurry”-multiple cut-and-paste errors

○ Poor statistical analysis of data

- Line listings not included
- Discordant analysis- check new statistical guidance

Why ? (2)

- Administrative gaps- missing documents

- Copies of IRB approval letters, IC ,financial disclosure forms, list of investigators....
- Clinical registration trial form, names and location of clinical sites....

- Lack of monitoring/auditing of clinical sites

Approval delayed by BIMO inspection findings

Why ? (3)

- Literature to support device-
Not analyzed appropriately, not summarized, organized
- Lack of knowledge about the clinical disease state -
end user Focus Panels!
- The “Intended Use” is the driving force of the review. Claim- supporting studies not adequate
- Issues with Quality System Inspection of
manufacturing facility. Poorly written manufacturing sections

Specific Software/Hardware Section (1)

- Hardware:

- Differences between clinical and launch platform not shown. Use of prototype for clinical trial not justified

- Claims for use needed for multiple amplification /detection platforms

- Assays need to be validated and cleared for each platform

- RUO labeled platform issue has prevented approval

Software/Hardware Section (2)

- Software :

- Guidance Document not followed

- <http://www.fda.gov/cdrh/ode/guidance/337.pdf>

- Summary of validation/verification testing not sufficient

- Need to link test results back to functional requirements
& link hazard analysis mitigations back to
functional requirements

- "Off the Shelf" software not sufficiently documented

- Guidance for Off-the-Shelf Software Use in Medical Devices*

- Minor "bugs" at launch? Justify why not a hazard and
mitigate through labeling

Device Design Section

Reagents

- Serological assays

- Did you characterize well antibodies/antigens?

- Nucleic acid assays

- primer/probe design justification required
 - include blast search results demonstrating specificity & inclusivity

- Include detailed description of appropriate internal and external controls/calibrators

Analytical/Clinical Study Sections (1)

- Precision/Reproducibility- minimum of 3 sites
 - Do panels assess variability of the assay at the cutoff/LOD?
- Samples/Populations/Sites
 - Do they represent the “Intended Use” population/end user?
- Non-US Patient Data-appropriate or not?
 - Check with FDA first

Analytical/Clinical Study Sections (2)

○ Specimen Type

Did you supply full analytical and clinical validation data to support claims for:

- Each specimen type
- All matrices
- All specimen collection devices
- All transport media
- All transport and storage conditions
- All collection methods

Analytical/Clinical Study Sections (3)

All NAAT assay extraction methods

- Should be validated with your assay

If “required but not part of kit”, check its regulatory status

RUO labeling of “ancillary reagents” has been an issue preventing device approval



What is Available to Streamline the Process?

Advice/Guidance Documents

FDA Pre-IDE Consultation

- Face-to-Face meetings
- Telecons

Interactive Submission Reviews



Pre-IDE Process


- Free FDA consult
- Protocol review and regulatory guidance
- Unique interactive opportunity (Non-binding)
- Especially recommended for novel devices/uses

<http://www.fda.gov/cdrh/oivd/presentations/042203-Altaie.html>

Information: CDRH Homepage

www.fda.gov/cdrh

- **Device Classification Database**
- **Device Advice**
 - <http://www.fda.gov/cdrh/devadvice>
- **Register for “What’s New”**
- **Guidance Documents**
- **IDE Information**
 - <http://www.fda.gov/cdrh/devadvice/ide/index.shtml>
- **Much more...**



Device Advice

Device Advice - Microsoft Internet Explorer

File Edit View Favorites Tools Help

Back Forward Stop Home Search Favorites Reload Print Mail News RSS Feeds

Address <http://www.fda.gov/cdrh/devadvice/> Go

FDA U.S. Food and Drug Administration Department of Health and Human Services

CENTER FOR DEVICES AND RADIOLOGICAL HEALTH

[FDA Home Page](#) | [CDRH Home Page](#) | [Search](#) | [A-Z Index](#)

DEVICE ADVICE Site Index Help

Device Advice Home CDRH Home Comments

Device Advice is CDRH's self-service site for medical device and radiation emitting product information. Device Advice is an interactive system for obtaining information concerning medical devices.

Search for in Device Advice Powered by Google

Guidance Documents	Overview of Regulations	Investigational Device Exemptions (IDE)
CDRH Databases	Is Your Product Regulated?	Premarket Approval
Code of Federal Regulations	Classify Your Device	Quality Systems
Regulatory Manuals	How to Market Your Device	Medical Device Labeling
International Information	Does Your Product Emit Radiation?	Medical Device Reporting
Consumer Information	Registering Your Establishment	Medical Device Recalls
	Listing Your Device	Importing Devices Exporting Devices
	Premarket Notification 510(k)	Medical Device Tracking
	510(k)/GMP Exemption	Postmarket Surveillance Studies

[CDRH Home Page](#) | [CDRH A-Z Index](#) | [Contact CDRH](#) | [Accessibility](#) | [Disclaimer](#)
[FDA Home Page](#) | [Search FDA Site](#) | [FDA A-Z Index](#) | [Contact FDA](#) | [HHS Home Page](#)

Center for Devices and Radiological Health / CDRH

Guidance Documents

Draft Guidance for Industry and FDA Staff

Establishing the Performance Characteristics of In Vitro Diagnostic Devices for the Detection or Detection and Differentiation of Influenza Viruses

DRAFT GUIDANCE

This guidance document is being distributed for comment purposes only.
Document issued on: [release date of FR Notice]

Comments and suggestions regarding this draft document should be submitted within 90 days of publication in the *Federal Register* of the notice announcing the availability of the draft guidance. Submit written comments to the Division of Dockets Management (HFA-305), Food and Drug Administration, 5630 Fishers Lane, rm. 1061, Rockville, MD 20852. Alternatively, electronic comments may be submitted to <http://www.fda.gov/dockets/ecomments>. All comments should be identified with the docket number listed in the notice of availability that publishes in the *Federal Register*.

For questions regarding this document contact Sally Hojvat at 240-276-0711 or by email at sally.hojvat@fda.hhs.gov.



U.S. Department of Health and Human Services
Food and Drug Administration
Center for Devices and Radiological Health
Office of In Vitro Diagnostic Device Evaluation and Safety
Division of Microbiology Devices

Guidance for Industry and FDA Staff

Interactive Review for Medical Device Submissions: 510(k)s, Original PMAs, PMA Supplements, Original BLAs, and BLA Supplements

Document Issued on: [release date as stated in FR Notice]

The information collection provisions in this guidance have been approved under OMB control number 0910-xxxx. This approval expires ????. An agency may not conduct, or sponsor and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number." (OMB Nos. and expiration dates are available at this site: <http://www.fda.gov/oc/omb/0910-xxxx/0910-xxxx.html> Please contact the Regulations Staff, if you do not know the appropriate approval number or expiration date).

For questions regarding this document, contact the Premarket Notification (510(k)) Section or the Premarket Approval Section of CDRH at 240-276-4040 or Leonard Wilson of CBER by phone at 301-827-0373 or by email at leonard.wilson@fda.hhs.gov.



U.S. Department of Health and Human Services
Food and Drug Administration
Center for Devices and Radiological Health
Center for Biologics Evaluation and Research

Other Related Guidances

- FDA and Industry Actions on Premarket Approval Applications (PMAs): Effect on FDA Review Clock and Goals, June 30, 2008
<http://www.fda.gov/cdrh/mdufma/guidance/1218.html#2a>
- Interactive Review for Medical Device Submissions: 510(k)s, Original PMAs, PMA Supplements, Original BLAs, and BLA Supplements, December 28, 2007
<http://www.fda.gov/cdrh/ode/guidance/1655.html>
- Real-Time Premarket Approval Application (PMA) Supplements
<http://www.fda.gov/cdrh/ode/guidance/673.html>
- Premarket Approval Application Modular Review, November 3, 2003
<http://www.fda.gov/cdrh/mdufma/guidance/835.html>
- Premarket Approval Application Filing Review, May 1, 2003 - <http://www.fda.gov/cdrh/ode/guidance/297.html>
- Post-Approval Studies –
http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfPMA/pma_pas.cfm

Transparency, Information on Web

[New Search](#)

[Back To Search Results](#)

510(k) Premarket Notification Database

Device Classification Name	Classifier, Prognostic, Recurrence Risk Assessment, Rna Gene Expression, Breast Cancer
510(K) Number	K062694
Device Name	MAMMAPRINT
Applicant	AGENDIA BV Louwesweg 6 Amsterdam, NL 1066 EC
Contact	Guido Brink
Regulation Number	866.6040
Classification Product Code	NYI
Date Received	09/11/2006
Decision Date	02/06/2007
Decision	Cleared For Marketing Automatic Class Iii Designat (AN)
Classification Advisory Committee	Immunology
Review Advisory Committee	Immunology
FOI ITEM	LETTER
FDA Review	Decision Summary
Type	Cleared For Marketing Automatic Class III Designation

FOI ITEM LETTER



DEPARTMENT OF HEALTH & HUMAN SERVICES

FEB - 6 2007

Agendia BV
c/o Mr. Guido Brink
Director Quality Management & Regulatory Affairs
Slotervaart Hospital, Floor 9D
Louwesweg 6, 1066 EC Amsterdam
The Netherlands

Re: k062694
Evaluation of Automatic Class III Designation
MammaPrint®
Regulation Number: 21 CFR 866.6040
Classification: Class II
Product Code: NYI

Dear Mr. Brink:

The Center for Devices and Radiological Health (CDRH) of (FDA) has completed its review of your petition for classification intended as a qualitative *in vitro* diagnostic test service, per

FDA Review Decision Summary (for a 510(k); SSED for a PMA)

510(k) SUBSTANTIAL EQUIVALENCE DETERMINATION DECISION SUMMARY

- A. 510(k) Number:
k062694
- B. Purpose for Submission:
New device
- C. Measurand:
70 gene expression profile
- D. Type of Test:
Expression microarray
Test service performed in a single laboratory in Agendia's Amsterdam facility.
- E. Applicant:
Agendia BV
- F. Proprietary and Established Names:
MammaPrint®
- G. Regulatory Information:
1. Regulation section:
21 CFR 866.6040 Gene expression profiling test system for breast cancer prognosis
 2. Classification:
Class II
 3. Product code:
NYI, Classifier, prognostic, recurrence risk assessment, RNA gene expression, breast cancer
 4. Panel:
Immunology (82)
- H. Intended Use:
1. Intended use(s):
MammaPrint® is a qualitative *in vitro* diagnostic test service, performed in a single laboratory, using the gene expression profile of fresh frozen breast cancer tissue samples to assess a patients' risk for distant metastasis.

The test is performed for breast cancer patients who are less than 61 years old, with Stage I or Stage II disease, with tumor size ≤ 5.0 cm and who are lymph node negative. The MammaPrint® result is indicated for use by physicians as a prognostic marker only, along with other clinicopathological factors.
 2. Indication(s) for use:
Same as intended use
 3. Special conditions for use statement(s):
For prescription use only
MammaPrint® is not intended for diagnosis, or to predict or detect response to therapy, or to help select the optimal therapy for patients.
 4. Special instrument requirements:
Agilent 2100 Bioanalyzer: Serial number DE54700497 en DE24802382
Agilent DNA microarray scanner: Serial number us22502555



Summary :Keys to a Successful PMA Submission

- Scientifically designed and well executed studies
- Good manufacturing practice documentation
- Appropriate statistical analysis of data
- Well written submission based on scientific principles
- Make use of available FDA documents and resources on the web
- Good communication with FDA throughout the entire process; pre-IDE meetings highly recommended



Questions ?
